

## **Certificate of Conformity No 213-01899**

Object	<b>Three-Phase Power Quality Measurement Device</b> Type: LINAX PQ5000 Transformer operated meter, accuracy class 0,2 S
Applicant	<b>CAMILLE BAUER METRAWATT AG</b> Aargauerstrasse 7 5610 Wohlen
Requirements	Partial conformity assessment according to the standard IEC 62053-22, subclauses: Limits of error due to variation of the current Starting current Meter constant Reversed phase sequence Harmonic components Sub-harmonics in the a.c. current circuit
Confirmation	Compliant with requirements
Date of Examination	7 to 14 January 2020
3003 Bern-Wabern, 13 March 2020	
For the Examination	Christian Santschi

Approved byDr Cédric Blaser, Head of LaboratoryLaboratory Electrical Energy and Power

This document is only valid and reviewable in its electronic form. Please observe the information given on <u>www.metas.ch/ecert</u>.